



Glossary



Manufacturer

Description

To identify the manufacturer of the product. This symbol is to be used in all applications to differentiate it from ISO 7000-2497.

Standard Title

Medical devices - The symbol is to be used with medical device labels, labelling and information to be supplied - Part 1: General requirement.

Referenced in

ISO 15223-1

ISO 9687:2015



Operations Manual

Description

To identify the locations where the operator's manual is sorted or to identify information that relates to the operating instructions. To indicate that the operating instructions should be considered when operating the device or control close to where the symbol is placed.

Standard Title

Medical devices – Symbols to be used with medical device labels, labelling and information to be supplied – Part 1: General requirements

Referenced in

ISO 5232

ISO/IEC 13251

ISO13998

ISO13999-1

ISO 11393-5

ISO 15223-1



Recycling

Description

To indicate that the package is recyclable.

Standard Title

Graphical symbols for use on equipment – Registered symbols

Referenced in

ISO 704

ISO/IEC 13251

ISO 10987-1

ISO 9687



Do not use if package is damaged

Description

To indicate that the device must not be used if the package or the device is damaged.

Standard Title

Medical devices – Symbols to be used with medical device labels, labelling and information to be supplied – Part 1: General requirements

Referenced in

ISO 15223-1



QR Code

Description

To direct users to Corneal Lens Corporation NZ Ltd's website

Standard Title

Information technology – Automatic identification and data capture techniques – QR Code 2005 bar code symbology specification

Referenced in

ISO/IEC 18004:2006

LOT **Batch Code****Description**

To identify the manufacturer's batch code, for example on a medical device or the corresponding packaging. The code shall be placed adjacent to the symbol

Standard Title

Medical devices – Symbols to be used with medical device labels, labelling and information to be supplied – Part 1: General requirement

Referenced in

ISO 15223-1

ISO 21531

ISO 9687

ISO 80601-2-55

ISO 80601-2-12

ISO 80601-2-67

ISO 80601-2-69

ISO 80601-2-70

REF **Catalogue Number****Description**

To identify the manufacturer's catalogue number. The catalogue number shall be placed adjacent to the symbol.

Standard Title

Medical devices – Symbols to be used with medical device labels, labelling and information to be supplied – Part 1: General requirement

Referenced in

ISO 15223-1

ISO 21531

ISO 9687

ISO 80601-2-55

ISO 80601-2-12

ISO 80601-2-67

ISO 80601-2-69

ISO 80601-2-70



Use By Date

Description

To indicate that the device should not be used after the date accompanying the symbol.

Standard Title

Medical devices – Symbols to be used with medical device labels, labelling and information to be supplied – Part 1: General requirement

Referenced in

ISO 15223-1

ISO 21531

ISO 9687

ISO 80601-2-55

ISO 80601-2-12

ISO 21420



Sterilised Using Steam Or Dry Heat

Description

To indicate that the device is provided sterile and has been sterilised using steam or dry heat.

Standard Title

Medical devices – Symbols to be used with medical device labels, labelling and information to be supplied – Part 1: General requirement

Referenced in

ISO 15223-1

ISO 21531

ISO 9687

ISO 80601-2-55

ISO 80601-2-12